

# The use of ultrasound for the estimation of muscle mass: one site fits most?

During the last two decades, DXA-derived appendicular lean soft tissue mass has served as a major criterion for diagnosing the age-related loss of skeletal muscle mass (i.e. sarcopenia).<sup>1</sup> From a clinical perspective, however, DXA measurements are costly, and the device has limited availability. A possible alternative method for measuring muscle size (e.g. muscle thickness) in humans may be B-mode ultrasound. It is well known that ultrasound is a non-invasive, quick, valid, and reliable imaging technique to estimate muscle thickness in muscles of the extremity and trunk.<sup>2,3</sup>

With this in mind, we read with great interest the article by Nijholt and colleagues<sup>4</sup> where they reported the validity and reliability of ultrasound to quantify musculature in older adults. The authors also reported on two prediction equations for estimating DXA-derived muscle mass. Although not reported within this paper, we previously noted that two of those prediction equations selected by Nijholt et al. included systematic error.<sup>5</sup> Over the last couple of years, we have published several prediction equations for estimating DXA-derived appendicular lean mass in older adults.<sup>6–8</sup> Unfortunately, those equations were not included in the article by Nijholt et al.<sup>4</sup> Interestingly, a single site measurement of forearm muscle thickness was found to be good predictor of DXA-derived lean soft tissue mass in older Caucasian adults,<sup>6</sup> and the equation was also found to be accurate in older Japanese adults.<sup>7</sup> Notably, the standard error of the estimate was 1.95 kg for the equation that used a single muscle thickness site, whereas the standard error of the estimate of the equation that included eight predictors was 1.13 kg.<sup>6</sup> The amount of time required for a single ultrasound measurement is generally less than 1 min per person, so this estimate appears both valid and pragmatic.

In summary, our previous studies<sup>6,7</sup> suggest that forearm muscle thickness measurements are a tolerable and less demanding assessment to use for older adults, and ultrasound estimated appendicular lean mass from the forearm muscle thickness may be a useful indicator for evaluating muscularity in older adults. Although additional research is needed, our recent work along with others noted within the Nijholt et al.<sup>4</sup> review may be useful with the development of ultrasound evaluation for health screenings as well as for the primary diagnosis of sarcopenia.

## Acknowledgements

The authors certify that they comply with the ethical guidelines for authorship and publishing of the *Journal of Cachexia, Sarcopenia and Muscle*.<sup>9</sup>

Takashi Abe

*Kevser Ermin Applied Physiology Laboratory, Department of Health, Exercise Science, & Recreation Management, The University of Mississippi, University, MS 38677, USA*  
t12abe@gmail.com

Jeremy P. Loenneke

*Kevser Ermin Applied Physiology Laboratory, Department of Health, Exercise Science, & Recreation Management, The University of Mississippi, University, MS 38677, USA*

Robert S. Thiebaud

*Department of Kinesiology, Texas Wesleyan University, 1201 Wesleyan St, Fort Worth, TX 76105, USA*

## References

1. Cruz-Jentoft AJ, Baeyens JP, Bauer JM, Boirie Y, Cederholm T, Landi F, et al. Sarcopenia: European consensus on definition and diagnosis: Report of the European Working Group on Sarcopenia in Older People. *Age Ageing* 2010;**39**:412–423.

2. Abe T, Loenneke JP, Thiebaud RS, Loftin M. Morphological and functional relationships with ultrasound measured muscle thickness of the upper extremity and trunk. *Ultrasound* 2014;**22**:229–235.
3. Abe T, Loenneke JP, Thiebaud RS. Morphological and functional relationships with ultrasound measured muscle thickness of the lower extremity: A brief review. *Ultrasound* 2015;**23**:166–173.
4. Nijholt W, Scafoglieri A, Jager-Wittenaar H, Hobbelen JS, van der Schans CP. The reliability and validity of ultrasound to quantify muscles in older adults: A systematic review. *J Cachexia Sarcopenia Muscle* 2017;**8**:702–712.
5. Abe T, Loenneke JP, Young KC, Thiebaud RS, Nahar VK, Hollaway KM, et al. Validity of ultrasound prediction equations for total and regional muscularity in middle-aged and older men and women. *Ultrasound Med Biol* 2015;**41**:557–564.
6. Abe T, Thiebaud RS, Loenneke JP, Young KC. Prediction and validation of DXA-derived appendicular lean soft tissue mass by ultrasound in older adults. *Age (Dordr)* 2015;**37**:114.
7. Abe T, Fujita E, Thiebaud RS, Loenneke JP, Akamine T. Ultrasound-derived forearm muscle thickness is a powerful predictor for estimating DXA-derived appendicular lean mass in Japanese older adults. *Ultrasound Med Biol* 2016;**42**:2341–2344.
8. Abe T, Loenneke JP, Thiebaud RS, Fujita E, Akamine T, Loftin M. Prediction and validation of DXA-derived appendicular fat-free adipose tissue by a single ultrasound image of the forearm in Japanese older adults. *J Ultrasound Med* 2017; <https://doi.org/10.1002/jum.14343>.
9. von Haehling S, Morley JE, Coats AJS, Anker SD. Ethical guidelines for publishing in the Journal of Cachexia, Sarcopenia and Muscle: update 2015. *J Cachexia Sarcopenia Muscle* 2015;**6**:315–316.